

Certificate of Analysis

Certificate of Analysis

Information

This certificate of analysis is only valid for the stated Lot Number and until the expiry date.

We hereby confirm with to the - mentioned product:

Lot no. 00003417

Expiry Date 07/2026

Description:

acid with detergent, applicable in laboratories for analytical chemistry and clinical diagnostics.

Packaging presentation:

Single Use Unit: 1 x 500 ml filled into non food toxic plastic bottle

Customary Unit: 1 x Unit secondary Packaging - Carton

Storage and Stability:

+15°C +25°C

Store contamination-free, well closed, at well tempered place, protected from solar radiation.

Parameter	Specification	Result	Unit
Conductivity κ (25.0 ± 0.02 °C) *1,3)	39.0 ± 2.0	complies	mS/cm
pH (20.0 °C) *1,3) =	1.10 ± 0.5	complies	
Foaming *2)	foaming	complies	
Appearance *2)	clear	complies	
Color *2)	colorless	complies	

*1) Measurement.

*2) Visual control.

*3) Traceable to primary standards (e.g. NIST, PTB, gravimetry etc.)

The product meets TECO's quality acceptance criteria.

D-84088 Neufahrn i.NB, 09.08.2023



Britta Steinert

Head of Quality Assurance

