

Certificate of Analysis

Lot No.: 012/2

Expiry Date: 10/2025

Description:

Stabilised system water, highest purity for analysis in analytical chemistry and clinical diagnostics laboratories.

Packaging presentation:

Single Use Unit: 1 x 1250 ml filled into non food toxic plastic bottle

Customary Unit: 3 x Single Use Unit in Carton

Storage and Stability:

+4°C +25°C

Store contamination-free, well closed, at well tempered place, protected from solar radiation.

PARAMETERS	SPECIFICATION	APPLIED STANDARD	PASS
<u>ph value</u>	7.0 ($\pm 1,0$ at 37 °C ± 1 °C)	DIN EN 27888	☒
elect. conductivity	< 0,1 μ S/cm (at 25°C)	DIN 38404 C5	☒
appearance :	Clear, colourless liquid	no precipitate or floting particulates	☒
CAS No.	[7732-18-5]	Chemical Abstracts Service	☒

The product meets TECO's quality acceptance criteria.
D-84088 Neufahrn 2023-01-09



Britta Steinert
Head of Quality Assurance

