

# Rinse solution

LAB

REF

21 500 00

## Certificate of Analysis

Lot No.: 015

Expiry Date: 08/2026

### Description:

Stabilised system water, highest purity for analysis in analytical chemistry and clinical diagnostics laboratories.

### Packaging presentation:

Single Use Unit: 1 x 1250 ml filled into non food toxic plastic bottle  
Customary Unit: 3 x Single Use Unit in Carton

### Storage and Stability:

+4°C .... +25°C

Store contamination-free, well closed, at well tempered place, protected from solar radiation.

| PARAMETERS          | SPECIFICATION                        | APPLIED STANDARD                       | PASS                                |
|---------------------|--------------------------------------|----------------------------------------|-------------------------------------|
| ph value            | 7.0 ( $\pm 1,0$ at 37 °C $\pm 1$ °C) | DIN 38404 C5                           | <input checked="" type="checkbox"/> |
| elect. conductivity | < 0,1 $\mu$ S/cm (at 25°C)           | DIN EN 27888                           | <input checked="" type="checkbox"/> |
| appearance :        | Clear, colourless liquid             | no precipitate or floting particulates | <input checked="" type="checkbox"/> |
| CAS No.             | [7732-18-5]                          | Chemical Abstracts Service             | <input checked="" type="checkbox"/> |

The product meets TECO's quality acceptance criteria.  
D-84088 Neufahrn 2023-08-23



Britta Steinert  
Head of Quality Assurance

