



MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFABRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis
Calcium Chloride 0.025M

REF

A0350-050

LOT

10353473



Mar.2025

Quality Control Specifications	Specification	Results	Pass
Appearance	Clear, colorless liquid	Clear, colorless liquid	✓
Odour	no characteristic odour	no characteristic odour	✓
pH	7,0 ($\pm 0,50$ at $37^{\circ}\text{C} \pm 1$)	6,99	✓

En: These values were determined by the laboratory of manufacturer according to their internal Quality Control Procedures. However, variations in analyzer, reagents, controls, technique etc., may cause different results. In this case the laboratory must establish its own control range according or refer to the certificate of control.

De: Diese Werte wurden durch das Labor von Hersteller gemäß der internen Qualitätsvorschrift ermittelt. Allerdings kann es je nach Gerät, Reagenzien, Kontrollen, Technik usw. vorkommen, dass die Werte unterschiedlich ausfallen. In diesem Fall muss das Labor den Kontrollbereich selbst bestimmen oder die Angaben des Analysezertifikats der Kontrolle verwenden. *