



**MEDICAL INSTRUMENTS**  
**PRODUCTION + TRADING GMBH**  
Internet: [www.teco-gmbh.com](http://www.teco-gmbh.com)  
Email: [info@teco-gmbh.com](mailto:info@teco-gmbh.com)

**DIESELSTRASSE 1**  
**D-84088 NEUFABRN NB.**  
Tel: #49-8773-70780-0  
Fax: #49-8773-70780-29

**Certificate of Analysis**  
**Calcium Chloride 0.025M**

**REF**

A0350-100

**LOT**

10353481



Jan. 2026

| Quality Control Specifications | Specification                                    | Results                 | Pass |
|--------------------------------|--------------------------------------------------|-------------------------|------|
| Appearance                     | Clear, colorless liquid                          | Clear, colorless liquid | ✓    |
| Odour                          | no characteristic odour                          | no characteristic odour | ✓    |
| pH                             | 7,0 ( $\pm 0,50$ at $37^{\circ}\text{C} \pm 1$ ) | 6,61                    | ✓    |

**En:** These values were determined by the laboratory of manufacturer according to their internal Quality Control Procedures. However, variations in analyzer, reagents, controls, technique etc., may cause different results. In this case the laboratory must establish its own control range according or refer to the certificate of control.

**De:** Diese Werte wurden durch das Labor von Hersteller gemäß der internen Qualitätsvorschrift ermittelt. Allerdings kann es je nach Gerät, Reagenzien, Kontrollen, Technik usw. vorkommen, dass die Werte unterschiedlich ausfallen. In diesem Fall muss das Labor den Kontrollbereich selbst bestimmen oder die Angaben des Analysezertifikats der Kontrolle verwenden. \*