



**MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH**
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

**DIESELSTRASSE 1
D-84088 NEUFAHRN NB.**
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

IBS Buffer

REF

A0590-125

LOT

10593315



May.2025

Quality Control Specifications	Specification	Result
Appearance	Clear colorless solution free of particulates	Pass
pH	7,2 - 7,6	7,2

En: These values were determined by the laboratory of Manufacturer according to their internal IVD Quality Control Procedures. However, variations in analyzer, reagents, controls, technique etc., may cause different results. In this case the laboratory must establish its own control range according or refer to the certificate of control.

De: Diese Werte wurden durch das Labor von Hersteller gemäß der internen IVD Qualitätsvorschrift ermittelt. Allerdings kann es je nach Gerät, Reagenzien, Kontrollen, Technik usw. vorkommen, dass die Werte unterschiedlich ausfallen. In diesem Fall muss das Labor den Kontrollbereich selbst bestimmen oder die Angaben des Analysezertifikats der Kontrolle verwenden. *

28.06.2022 Teco QM department

WG

Digitally signed by WG
DN: cn=WG, o=TECO GmbH,
ou=Quality Control Department,
email=info@teco-medical.de, c=DE