



**MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH**
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

**DIESELSTRASSE 1
D-84088 NEUFAHRN NB.**
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

Owren's Buffer

REF

A0592-100

LOT

10593879



Sep. 2027

Quality Control Specifications	Specification	Result
Appearance	Clear solution	Approved
pH	7,2 - 7,4	7,3
Clotting time normal control	Within $\pm 1,5$ s from analysis wiht reference buffer	-0,4 s
Clotting time abnormal control	Within $\pm 2,5$ s from analysis wiht reference buffer	-1,6 s

Quality Control Department

REV-1