



MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFAHRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

Dilution Buffer

REF

A0593-035

LOT

32023534



Jun. 2026

Analyses	Specification	Result
pH	7,2 - 7,4	7,3
Conductivity	17 - 21 mS/cm	18 mS/cm