

Certificate of Analysis

Certificate of Analysis

Information

This certificate of analysis is only valid for the stated Lot Number and until the expiry date.
In accordance with the applicable directives and existing operating requirements,
we hereby confirm with respect to the above-mentioned product:

Lot no. 00003770
Expiry Date 01/2027

Parameter	Specification	Result	Unit
Conductivity κ (25.0 $\pm$ 0.02 °C) *1, 3)	39.0 $\pm$ 2.0	complies	mS/cm
pH (20.0 °C) *1, 3) =	1.10 $\pm$ 0.5	complies	
Foaming *2)	foaming	complies	
Appearance *2)	clear	complies	
Color *2)	colorless	complies	

*1) Measurement.

*2) Visual control.

*3) Traceable to primary standards (e.g. NIST, PTB, gravimetry etc.)

The product meets TECO's quality acceptance criteria for intended purpose.
D-84088 Neufahrn i.NB, 16.01.2025



Simone Spitzl
Head of Quality Assurance