MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFAHRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis
TEClot PT-S

(expected values for analysers of TECO Coatron series)

REF A0230-100 LOT 10233999 28.08.2029

STD (%)	Coatron X 1.04.52	Coatron A V1.07.05a	Coatron M4 C1.13d	KC-4 Micro
100,0 %	13,3 s	13,1 s	13,9 s	14,3 s
50,0 %	17,4 s	19,0 s	18,9 s	20,0 s
25,0 %	25,1 s	30,7 s	29,9 s	33,9 s
12,5 %	55,5 s	55,9 s	60,4 s	70,0 s

ISI value	1,17			
Normal time	13,3 s	13,1 s	13,9 s	14,3 s

Control Range	see certificate of control

En: These values were determined by the laboratory of ***TECO GmbH*** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration curve and follow local requirements and regulations for quality control.

De:Diese Werte wurden durch das Labor von ***TECO GmbH*** gemäß der Verfahrensanweisung QMV-07-26 ermittelt. Jedoch kann es je nach Charge der Reagenzien, IBS Puffer oder anderer Verdünnungslösungen, Kalibratoren, Kontrollen oder je nach Gerät, Pipettiertechnik usw. vorkommen dass die Werte von Labor zu Labor unterschiedlich ausfallen. Deshalb sind die angeführten Ergebnisse nur als Richtwerte zu bewerten. Jedes Labor muss die Eichkurve selbst bestimmen und die örtlichen Anforderungen für Qualitätskontrollen anwenden.

Signed by:
Wendy Guo
QC Department, TECO

2026.10.08 14:14:41+02'00'