MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFAHRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TEClot PT-B

(expected values for analysers of TECO Dimex Jr. & Coatron X Eco)

REF A0260-050 LOT 10263995 09.2027

Typical standard curve*	PT (sec)
100,0 %	23,4 s
25,0 %	61,5 s

(refer to 1:6 dilution procedure)

ISI	1,11
Normal	23,4 s
Control	see certificate of control

*

En: These values were determined by the laboratory of ***TECO GmbH*** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration curve and follow local requirements and regulations for quality control.

De:Diese Werte wurden durch das Labor von ***TECO GmbH*** gemäß der Verfahrensanweisung QMV-07-26 ermittelt. Jedoch kann es je nach Charge der Reagenzien, IBS Puffer oder anderer Verdünnungslösungen, Kalibratoren, Kontrollen oder je nach Gerät, Pipettiertechnik usw. vorkommen, dass die Werte von Labor zu Labor unterschiedlich ausfallen. Deshalb sind die angeführten Ergebnisse nur als Richtwerte zu bewerten. Jedes Labor muss die Eichkurve selbst bestimmen und die örtlichen Anforderungen für Qualitätskontrollen anwenden.

Signed by: Wendy Guo
QC Department, TECO

2026.10.05
09:06:43+02'00'