



MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFABRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis
Calcium Chloride 0.025M

REF A0350-050

LOT 10353732

 Sep. 2026

Quality Control Specifications	Specification	Results	Pass
Appearance	Clear, colorless liquid	Clear, colorless liquid	✓
Odour	no characteristic odour	no characteristic odour	✓
pH	7,0 ($\pm 0,50$ at $37^{\circ}\text{C} \pm 1$)	6,81	✓

En: These values were determined by the laboratory of **TECO GmbH** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration curve and follow local requirements and regulations for quality control.

De: Diese Werte wurden durch das Labor von **TECO GmbH** gemäß der Verfahrensanweisung QMV-07-26 ermittelt. Jedoch kann es je nach Charge der Reagenzien, IBS Puffer oder anderer Verdünnungslösungen, Kalibratoren, Kontrollen oder je nach Gerät, Pipettiertechnik usw. vorkommen dass die Werte von Labor zu Labor unterschiedlich ausfallen. Deshalb sind die angeführten Ergebnisse nur als Richtwerte zu bewerten. Jedes Labor muss die Eichkurve selbst bestimmen und die örtlichen Anforderungen für Qualitätskontrollen anwenden.