



**MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH**
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

**DIESELSTRASSE 1
D-84088 NEUFAHRN NB.**
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TEClot FIB

REF A0511-020

LOT 10513971

 2026-05-30

 2028-05-31

QC procedures	Requirement	Result	Status
---------------	-------------	--------	--------

Visual inspection

Appearance before recon.	white colored substrate	confirmed	ok
Appearance after recon.	clear liquid	confirmed	ok

Calibration inspection

Slope	-1,38 to -1,24	-1,37	ok
Linearity (R ²)	0,994 - 0,94	0,994	ok

Trueness inspection

TEControl N	LOT=96003797	238 - 336mg/dl	310mg/dl	ok
TEControl A	LOT=96103819	104 - 156mg/dl	140mg/dl	ok

Application notes

Each laboratory must establish its own calibration curve for each new lot of reagent and IBS buffer !

The IBS buffer may exhibit lot-to-lot variability of up to 1.5 seconds. Do not store different lots of IBS buffer together in your refrigerator and perform new calibration, whenever IBS buffer lot is changed.

Do not store or prewarm Fibrinogen reagent to 37°C after reconstitution.
The reagent can be applied cold (2-8°C) or at room temperature on Coatron X.