



**MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH**
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

**DIESELSTRASSE 1
D-84088 NEUFAHRN NB.**
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TEChrom AT liquid (anti-Xa)

(expected values for analysers of TECO Coatron series)

REF C1010-020

LOT 21013619

 Feb.2025

typical values ⁽¹⁾

AT (%)	Coatron M	Coatron X	Coatron A
92,0	0,354 E	0,339 E	0,035 E
46,0	0,542 E	0,589 E	0,175 E
23,0	0,613 E	0,722 E	0,275 E

⁽¹⁾There are no specific control target values assigned (required e.g. for Rilibaek). These values can be read in the certificate of analysis of control plasma. Instead only typical calibration data & control ranges by the laboratory of **TECO GmbH** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration curve and follow local requirements and regulations for quality control.