



MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFABRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TEChrom AT liquid (anti-Xa)

(expected values for analysers of TECO Coatron series)

REF

C1010-020

LOT

21013871



Jun.2027

typical values ⁽¹⁾

AT (%)	Coatron M	Coatron X	Coatron A
102,0	0,383 E	0,420 E	0,037 E
51,0	0,591 E	0,645 E	0,185 E
25,5	0,696 E	0,796 E	0,290 E

⁽¹⁾There are no specific control target values assigned (required e.g. for Rilibaek). These values can be read in the certificate of analysis of control plasma. Instead only typical calibration data & control ranges by the laboratory of **TECO GmbH** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration