



MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFAHRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TEChrom AT liquid (anti-Xa)

REF C1010-020

LOT 21013908

 Jun.2027

Components

Factor Xa Reagent 6mL
Factor Xa Substrate 3mL

Storage

Store at 2 - 8 °C

Conformity Statement

This product was manufactured in accordance with TECO's ISO 13485 certified quality management system and meets all of TECO's quality control specifications.

Analysis	Unit	Specification	Result
Substrate appearance	visual	Clear solution	approved
Substrate concentration	mg/mL	1.52 - 1.68 mg/mL	1.55 mg/mL

Calibration curve

Slope	AT(%) / dE	-75 to -135	-124
Intercept	AT(%)	100 - 180	142
Linearity	R ²	> 0.98	1.00