



**MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH**
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

**DIESELSTRASSE 1
D-84088 NEUFAHRN NB.**
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TEChrom AT liquid (anti-Xa)

(expected values for analysers of TECO Coatron series)



C1010-020



21013987



Jun.2028

Components

Factor Xa Reagent 6mL

Factor Xa Substrate 3mL

Storage

Store at 2 - 8 °C

Conformity Statement

This product was manufactured in accordance with TECO's ISO 13485 certified quality management system and meets all of TECO's quality control specifications.

Analysis	Unit	Specification	Result
Substrate appearance	visual	Clear solution	approved
Substrate concentration	mg/mL	1.52 - 1.68 mg/mL	1.53 mg/mL

Calibration curve

Slope	AT(%) / dE	-75 to -135	-126
Intercept	AT(%)	100 - 180	148
Linearity	R ²	> 0.99	1.00