



**MEDICAL INSTRUMENTS  
PRODUCTION + TRADING GMBH**  
Internet: [www.teco-gmbh.com](http://www.teco-gmbh.com)  
Email: [info@teco-gmbh.com](mailto:info@teco-gmbh.com)

**DIESELSTRASSE 1  
D-84088 NEUFAHRN NB.**  
Tel: #49-8773-70780-0  
Fax: #49-8773-70780-29

## Certificate of Analysis

### Factor II Deficient Plasma

**REF** P5001-010

**LOT** 95003798

 31.1.2025

| Analyses        |        | Specifications | Results  | Status |
|-----------------|--------|----------------|----------|--------|
| Factor II level |        | <3%            | 0,0%     | passed |
| pH              |        | 7,4 - 7,8      | 7,7      | passed |
| Virology        | HBsAg  | negative       | negative | passed |
|                 | HIV-Ab | negative       | negative | passed |
|                 | HCV-Ab | negative       | negative | passed |