



MEDICAL INSTRUMENTS  
PRODUCTION + TRADING GMBH  
Internet: [www.teco-gmbh.com](http://www.teco-gmbh.com)  
Email: [info@teco-gmbh.com](mailto:info@teco-gmbh.com)

DIESELSTRASSE 1  
D-84088 NEUFABRN NB.  
Tel: #49-8773-70780-0  
Fax: #49-8773-70780-29

## Certificate of Analysis

### Factor V Deficient Plasma

**REF** P5101-010

**LOT** 95103718

 Jun.2026

| Analyses       |        | Specifications | Results  | Status |
|----------------|--------|----------------|----------|--------|
| Factor V level |        | <3%            | 0,2%     | passed |
| pH             |        | 7,4 - 7,8      | 7,6      | passed |
| Virology       | HBsAg  | negative       | negative | passed |
|                | HIV-Ab | negative       | negative | passed |
|                | HCV-Ab | negative       | negative | passed |