



MEDICAL INSTRUMENTS  
PRODUCTION + TRADING GMBH  
Internet: [www.teco-gmbh.com](http://www.teco-gmbh.com)  
Email: [info@teco-gmbh.com](mailto:info@teco-gmbh.com)

DIESELSTRASSE 1  
D-84088 NEUFABRN NB.  
Tel: #49-8773-70780-0  
Fax: #49-8773-70780-29

## Certificate of Analysis

### Factor VIII Deficient Plasma

**REF**

P5301-010

**LOT**

95303588



May.2025

| Analyses          |        | Specifications | Results  | Status |
|-------------------|--------|----------------|----------|--------|
| Factor VIII level |        | <1%            | 0,2%     | passed |
| pH                |        | 7,4 - 7,8      | 7,5      | passed |
| Virology          | HBsAg  | negative       | negative | passed |
|                   | HIV-Ab | negative       | negative | passed |
|                   | HCV-Ab | negative       | negative | passed |

Quality Control Department

REV-1