



MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFABRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

Factor X Deficient Plasma

REF P5501-010

LOT 95503953

 Jan.2027

Analyses		Specifications	Results	Status
Factor X level		<3%	0,0%	passed
pH		7,4 - 7,8	7,5	passed
Virology	HBsAg	negative	negative	passed
	HIV-Ab	negative	negative	passed
	HCV-Ab	negative	negative	passed