



MEDICAL INSTRUMENTS  
PRODUCTION + TRADING GMBH  
Internet: www.teco-gmbh.com  
Email: info@teco-gmbh.com

DIESELSTRASSE 1  
D-84088 NEUFARN NB.  
Tel: #49-8773-70780-0  
Fax: #49-8773-70780-29

## Certificate of Analysis

### TEControl N

REF P6001-010

LOT 96003697



Sep.2026

| Prothrombin Time (PT) |                           |      |        |       |        |
|-----------------------|---------------------------|------|--------|-------|--------|
| Reagent               | Method                    | Unit | Target | Range |        |
| TEClot PT-S           | Coatron M                 | sec  | 14,2   | 11,6  | - 16,7 |
|                       |                           | %    | 99     | 81    | - 117  |
|                       |                           | INR  | 1,01   | 0,83  | - 1,19 |
|                       | Coatron A                 | sec  | 13,0   | 10,7  | - 15,3 |
|                       |                           | %    | 104    | 85    | - 122  |
|                       |                           | INR  | 0,98   | 0,80  | - 1,15 |
|                       | Coatron X                 | sec  | 13,0   | 10,7  | - 15,3 |
|                       |                           | %    | 102    | 84    | - 121  |
|                       |                           | INR  | 0,99   | 0,81  | - 1,17 |
|                       | KC4 micro                 | sec  | 14,2   | 11,6  | - 16,8 |
|                       |                           | %    | 100    | 82    | - 118  |
|                       |                           | INR  | 1,00   | 0,82  | - 1,18 |
| TEClot PT-B           | Coatron X Eco + Dimex Jr. | sec  | 22,5   | 18,5  | - 26,6 |
|                       |                           | %    | 104    | 85    | - 123  |
|                       |                           | INR  | 0,97   | 0,80  | - 1,15 |

| Activated Partial Thromboplastin Time (aPTT) |           |     |      |      |        |
|----------------------------------------------|-----------|-----|------|------|--------|
| TEClot APTT-S                                | Coatron M | sec | 33,0 | 27,0 | - 38,9 |
|                                              | Coatron A | sec | 31,6 | 25,9 | - 37,2 |
|                                              | Coatron X | sec | 31,2 | 25,5 | - 36,8 |
|                                              | KC4 micro | sec | 31,2 | 25,6 | - 36,8 |

| Fibrinogen (method Clauss, diluted 1:10) |           |       |     |     |       |
|------------------------------------------|-----------|-------|-----|-----|-------|
| TEClot FIB                               | Coatron M | mg/dL | 293 | 228 | - 357 |
|                                          | Coatron X | mg/dL | 272 | 212 | - 332 |
|                                          | Coatron A | mg/dL | 293 | 228 | - 357 |

| Thrombin Time (TT) |               |     |      |    |      |
|--------------------|---------------|-----|------|----|------|
| TEClot TT          | Coatron M + X | sec | 19,4 | 16 | - 23 |
|                    | Coatron A     | sec | 20,5 | 17 | - 24 |

| Antithrombin |           |   |     |    |       |
|--------------|-----------|---|-----|----|-------|
| Techrom AT   | chromogen | % | 106 | 83 | - 129 |

| D-Dimer (DD)    |                  |       |      |     |        |
|-----------------|------------------|-------|------|-----|--------|
| Blue D-Dimer LC | latex immuno DDU | ng/mL | 312  | 234 | - 390  |
|                 | latex immuno FEU | ng/mL | 780  | 639 | - 920  |
| RED D-Dimer     | latex immuno DDU | ng/mL | 420  | 315 | - 525  |
|                 | latex immuno FEU | ng/mL | 1051 | 788 | - 1313 |

These values were determined by the laboratory of **TECO GmbH** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration curve and follow local requirements and regulations for quality control.