



**MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH**
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

**DIESELSTRASSE 1
D-84088 NEUFAHRN NB.**
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TEControl N

REF P6001-010

LOT 96003902



Jun.2027

Prothrombin Time (PT)					
Reagent	Method	Unit	Target	Range	
TEClot PT-S	Coatron M	sec	14,8	12,1	- 17,5
		%	108	89	- 128
		INR	0,95	0,78	- 1,12
	Coatron A	sec	13,7	11,2	- 16,2
		%	103	84	- 122
		INR	0,98	0,81	- 1,16
	Coatron X	sec	13,8	11,3	- 16,3
		%	109	90	- 129
		INR	0,95	0,78	- 1,12
	KC4 micro	sec	14,9	12,2	- 17,6
		%	107	88	- 127
		INR	0,95	0,78	- 1,12
TEClot PT-B	Coatron X Eco + Dimex Jr.	sec	22,8	18,7	- 26,9
		%	98	81	- 116
		INR	1,01	0,83	- 1,19

Activated Partial Thromboplastin Time (aPTT)					
TEClot APTT-S	Coatron M	sec	31,5	25,8	- 37,1
	Coatron A	sec	30,2	24,8	- 35,6
	Coatron X	sec	30,0	24,6	- 35,3
	KC4 micro	sec	31,3	25,6	- 36,9

Fibrinogen (method Clauss, diluted 1:10)					
TEClot FIB	Coatron M	mg/dL	329	257	- 401
	Coatron X	mg/dL	261	204	- 319
	Coatron A	mg/dL	294	229	- 358

Thrombin Time (TT)					
TEClot TT	Coatron M + X	sec	19,5	16	- 23
	Coatron A	sec	21,5	18	- 25

Antithrombin					
Techrom AT	chromogen	%	100	78	- 122

D-Dimer (DD)					
Blue D-Dimer LC	latex immuno DDU	ng/mL	308	231	- 385
	latex immuno FEU	ng/mL	771	632	- 909
RED D-Dimer	latex immuno DDU	ng/mL	423	317	- 529
	latex immuno FEU	ng/mL	1058	793	- 1322

These values were determined by the laboratory of **TECO GmbH** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration curve and follow local requirements and regulations for quality control.