



**MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH**
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

**DIESELSTRASSE 1
D-84088 NEUFAHRN NB.**
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis TEControl A

REF P6101-010

LOT 96103819

Apr.2027

Prothrombin Time (PT)					
Reagent	Method	Unit	Target	Range	
TEClot PT-S	Coatron M	sec	26,9	21,3	- 32,6
		%	28	22	- 34
		INR	2,05	1,62	- 2,48
	Coatron A	sec	25,2	19,9	- 30,5
		%	27	22	- 33
		INR	2,11	1,67	- 2,55
	Coatron X	sec	25,2	19,9	- 30,5
		%	26	20	- 31
		INR	2,05	1,62	- 2,48
	KC4 micro	sec	28,3	22,4	- 34,2
		%	28	22	- 34
		INR	2,13	1,69	- 2,58
TEClot PT-B	Coatron X Eco + Dimex Jr.	sec	50,9	40,2	- 61,6
		%	28	22	- 34
		INR	2,46	1,95	- 2,98

Activated Partial Thromboplastin Time (aPTT)					
TEClot APTT-S	Coatron M	sec	61,9	48,9	- 74,9
	Coatron A	sec	66,0	52,1	- 79,8
	Coatron X	sec	59,0	46,6	- 71,3
	KC4 micro	sec	64,3	50,8	- 77,8

Fibrinogen (method Clauss, diluted 1:10)					
TEClot FIB	Coatron M	mg/dL	150	112	- 187
	Coatron X	mg/dL	146	110	- 183
	Coatron A	mg/dL	143	107	- 178

Thrombin Time (TT)					
TEClot TT	Coatron M + X	sec	n.a.	-	
	Coatron A	sec	n.a.	-	

Antithrombin					
Techrom AT	chromogen	%	49	37	- 61

D-Dimer (DD)					
Blue D-Dimer LC	latex immuno DDU	ng/mL	1073	697	- 1448
	latex immuno FEU	ng/mL	2682	2119	- 3246
RED D-Dimer	latex immuno DDU	ng/mL	1019	662	- 1376
	latex immuno FEU	ng/mL	2548	1656	- 3439

These values were determined by the laboratory of **TECO GmbH** according to standard operation procedure QMV-07-26. However, Lot-to-Lot variation of reagents, IBS buffer and other diluents, calibrators, controls as well as instruments or pipetting technique etc., may cause lab-to-lab different results. Therefore these assigned values must be seen as typical values. Each laboratory must establish its own calibration curve and follow local requirements and regulations for quality control.