



MEDICAL INSTRUMENTS
PRODUCTION + TRADING GMBH
Internet: www.teco-gmbh.com
Email: info@teco-gmbh.com

DIESELSTRASSE 1
D-84088 NEUFAHRN NB.
Tel: #49-8773-70780-0
Fax: #49-8773-70780-29

Certificate of Analysis

TECAL DD

REF

P8200-005

LOT

98203608



May.2025

Analyses	Specifications	Results	Status
Concentration DDU ¹	2880 - 3520 ng/mL	3320 ng/mL	passed
Concentration FEU ¹	7200 - 8800 ng/mL	8300 ng/mL	

(1) DDU = D-dimer units FEU = Fibrinogen Equivalent Units 1 FEU = 2.50 DDU