



**MEDICAL INSTRUMENTS  
PRODUCTION + TRADING GMBH**  
Internet: [www.teco-gmbh.com](http://www.teco-gmbh.com)  
Email: [info@teco-gmbh.com](mailto:info@teco-gmbh.com)

**DIESELSTRASSE 1  
D-84088 NEUFAHRN NB.**  
Tel: #49-8773-70780-0  
Fax: #49-8773-70780-29

## Certificate of Analysis

### TECAL DD

**REF**

P8200-005

**LOT**

98203823



Apr. 2027

| Analyses                       | Specifications    | Results    | Status |
|--------------------------------|-------------------|------------|--------|
| Concentration DDU <sup>1</sup> | 2880 - 3520 ng/mL | 2953 ng/mL | passed |
| Concentration FEU <sup>1</sup> | 7200 - 8800 ng/mL | 7383 ng/mL |        |

(1) DDU = D-dimer units    FEU = Fibrinogen Equivalent Units    1 FEU = 2.50 DDU